

# NDIR GAS ANALYZER

DATA SHEET

ZPA

## OVERVIEW

This gas analyzer (ZPA) can measure the concentration of NO, SO<sub>2</sub>, CO<sub>2</sub>, CO, CH<sub>4</sub> and O<sub>2</sub> in sample gas. NO, SO<sub>2</sub>, CO<sub>2</sub>, CO and CH<sub>4</sub> are measured by non-dispersion infrared method (NDIR), while O<sub>2</sub> is measured by a galvanic fuel cell, paramagnetic sensor, or zirconia sensor.

Up to five components including O<sub>2</sub> can be measured.

ZPA also features its compact design that enables downsizing of measurement system, and simple structure with single-beam IR system that requires minimal maintenance.

ZPA is best suited to the measurement of exhaust gas from refuse incinerators, boilers, and various industrial furnaces.



## FEATURES

### 1. Compact and lightweight

H133 × W483 × D382 mm, approximate 11 kg

### 2. Easy maintenance

Maintenance is easy due to the simple measurement unit of single-beam system adapted.

### 3. User-friendly operation

Clear and easy-to-read display of all five gas concentrations at once.

Simple status and maintenance messages.

### 4. Extensive functions

Various optional functions are available such as auto calibration control, atmospheric pressure correction, high and low concentration alarms, remote range switch, and range identification signal, etc.

- Max. 5 components measurement including O<sub>2</sub>.

For reverse range O<sub>2</sub> measurement, infrared gas measurement is not available; only the single range O<sub>2</sub> measurement is available.

- Measuring ranges are changeable between the specified minimum and maximum range
- Maximum rangeability. 1:10 (except O<sub>2</sub>)
- Up to two ranges are available for each component
- For possible combinations of components and ranges, refer to Table1.

### Measured value indication:

Digital indication in 4 digits

(LCD panel with LED back light)

- Instantaneous value of each component
- Instantaneous value after O<sub>2</sub> correction (only in NO, SO<sub>2</sub>, CO measurement with O<sub>2</sub>)
- Average value after O<sub>2</sub> correction (only in NO, SO<sub>2</sub>, CO measurement with O<sub>2</sub>)
- O<sub>2</sub> average value

### Analog output signals:

4 to 20mA DC or 0 to 1V DC, isolated internally from circuit and ground. Output lines are non-isolated each other.; 12 outputs max.

Allowable load 550Ω for 4 to 20mA DC

Allowable load 100KΩ for 0 to 1V DC

\* Refer to Table2 for the channel No. of displayed values and analog output signals.

### Analog input signal:

For signal input from external O<sub>2</sub> analyzer.

- (1) Signal from Fuji's Zirconia O<sub>2</sub> analyzer (TYPE: ZFK7)
- (2) 0 to 1V DC full-scale signal

\* Input section is not isolated.

\* External O<sub>2</sub> analyzer should be purchased separately.

### Digital output (Option):

1 form C contact (24V DC/1A, resistive load) Up to 15 outputs

Instrument error, calibration error, range identification, auto calibration status, solenoid valve drive for auto calibration, high/low limit alarm

\* All relay contacts are isolated mutually and from the internal circuit.

## SPECIFICATIONS

### Standard Specifications

#### Principle of measurement:

NO, SO<sub>2</sub>, CO<sub>2</sub>, CO, CH<sub>4</sub>;

Non-dispersion infrared-ray absorption method

Single light source and single beams (single beam system)

O<sub>2</sub>;

Galvanic fuel cell O<sub>2</sub> analyzer (built-in) paramagnetic O<sub>2</sub> analyzer (built-in), or zirconia O<sub>2</sub> analyzer (externally installed TYPE: ZFK7)

#### Measurable gas components and ranges:

	Minimum range	Maximum range
NO	0–200 ppm	0–5000 ppm
SO <sub>2</sub>	0–200 ppm	0–10 vol%
CO <sub>2</sub>	0–100 ppm	0–100 vol%
CO	0–200 ppm	0–100 vol%
CH <sub>4</sub>	0–500 ppm	0–100 vol%
O <sub>2</sub> (Built in fuel cell)	0–10 vol%	0–25 vol%
O <sub>2</sub> (Built in paramagnetic)	0–5 vol%	0–100 vol%
	None	100–95 vol%
O <sub>2</sub> (External zirconia)	0–5 vol%	0–25 vol%

**Digital input (Option):**

Voltage contact (12-24V DC, ≤15mA)

Up to 9 inputs

Remote range change over, auto calibration remote start, remote hold, average value reset.

\* Isolated from the internal circuit with photocoupler.

**Power supply:**

Voltage rating; 100V to 240V AC

Allowable range; 85V to 264V AC

Frequency; 50Hz/60Hz

Power consumption; 100VA max.

**Operating conditions:**

Ambient temperature; -5°C to 45°C

(40°C max. when using two optical systems with 200V AC power source)

Ambient humidity; 90% RH max., non-condensing

**Storage conditions:**

Ambient temperature; -20°C to 60°C

Ambient humidity; 95% RH max., non-condensing

**Dimensions (H × W × D):**

133 x 483 x 382mm

**Weight:**

Approx. 11 kg

**Finish color:**

Front panel; Cool gray (PANTON 1C-F)

**Enclosure:**

Steel casing, for indoor use

**Material of gas-contacting parts:**

Gas inlet/outlet; Stainless steel 304

Sample cell; Stainless steel 304, chloroprene rubber

Infrared-ray transmitting window; CaF<sub>2</sub>

Internal piping; vinyl chloride, PTFE, Polypropylene

Paramagnetic O<sub>2</sub> analyzer cell: Stainless steel 316

Fuel cell O<sub>2</sub> analyzer cell: ABS resin

**Gas inlet/outlet:**

Rc1/4 or NPT1/4 internal thread

**Purge gas flow rate:**

1L/min (when required)

**Life time of galvanic fuel cell O<sub>2</sub> analyzer:**

2 years

**Standard Functions****Output signal holding:**

Enables you to hold the output signal during calibration, to the value right before the calibration is started or the user-specified value. Values indicated on LCD will not be held.

**Range changeover:**

You can change between ranges by manually, automatically, or remotely.

Manual: by key operation

Auto: When the measured value reaches above 90% FS of the 1st range, the range automatically switches to the 2nd range. When the measured value goes down below 80% FS of the 1st range, the range automatically switches from the 2nd range to the 1st range.

Remote: by the contact input (option).

When the specified voltage (the remote range change-over signal) is applied on the contact dedicated for each component, the 1st range is effective. When no voltage is applied, the 2nd range becomes effective.

**Optional Functions****Remote output holding:**

Applying the specified voltage on the dedicated terminal allows you to hold the output signal to the last value or the user-specified value. Holding is effective while the voltage is applied. Values indicated on LCD are not held.

**Range identification signal:**

You can check which range is in use.

**Auto calibration:**

This function requires standard gas cylinders for calibration and solenoid valves for opening/closing the gas flow line. When this function is activated, the analyzer opens and closes the solenoid valve driving contact periodically at preset cycle.

Auto calibration cycle setting:

1 hour to 99 hours (in increments of 1 hour) or

1 day to 40 days (in increments of 1 day).

Gas flow time setting:

The time during which calibration gas is drawn

60 seconds to 900 seconds (in increments of 1 second)

**Auto calibration remote start:**

When you apply the specified voltage for 1.5 seconds or longer on the auto calibration remote start contact and then open the contact, one-time auto calibration starts. Calibration gas is drawn for the time set in the "gas flow time setting" for Auto calibration (see the previous item).

**Auto zero calibration:**

This function requires a standard gas cylinder for zero calibration and a solenoid valve for opening/closing the gas flow line. When this function is activated, the analyzer opens and closes the solenoid valve driving contact periodically at preset cycle. The cycle for the auto zero calibration and that for the auto calibration can be different.

Auto calibration cycle setting:

1 hour to 99 hours (in increments of 1 hour) or

1 day to 40 days (in increments of 1 day).

Gas flow time setting:

The time during which calibration gas is drawn

60 seconds to 900 seconds (in increments of 1 second)

**Upper/lower limit alarm:**

When an instantaneous value has gone beyond the upper limit or below the lower limit, the analyzer closes the contact to emit an alarm signal. Up to four alarms are available.

**Instrument error contact output:**

The contact is closed if a device error occurs.

**Calibration error contact output:**

The contact is closed if a calibration error occurs.

**Auto calibration status contact output:**

The contact is closed during auto calibration.

**O<sub>2</sub> correction:**

Conversion of measured NO, CO, and SO<sub>2</sub> gas concentrations into values at reference O<sub>2</sub> concentration

Correction formula:

$$C = \frac{21 - O_n}{21 - O_s} \times C_s$$

C: Sample gas concentration after O<sub>2</sub> correction

C<sub>s</sub>: Measured concentration of sample gas

O<sub>s</sub>: Measured O<sub>2</sub> concentration

O<sub>n</sub>: Reference O<sub>2</sub> concentration

(changeable by setting)

\* The upper limit value of the fractional part in this calculation is 4. The result of calculation is indicated and transmitted as an analog output signal.

### Average value after O<sub>2</sub> correction and O<sub>2</sub> average value calculation:

The analyzer can take measurement every 30 seconds, and calculate the moving average of instantaneous concentration after O<sub>2</sub> correction or instantaneous O<sub>2</sub> value per the period you set; in the range 1–59 min (in one minute increment) or 1–4 hour (in one hour increment). The analyzer transmits the moving average output every 30 seconds.

### Average value resetting:

The above-mentioned output of average value is started from the initial state by opening the average value resetting input terminals after short circuiting for 1.5 sec or longer. Output is reset by input voltage and restarted by opening the terminal circuit.

### Communication function:

RS-485 (9pins D-sub connector)  
 Half-duplex bit serial  
 Start-stop synchronization  
 Modbus RTU™ protocol  
 Contents: Read/Write parameters  
     Read measurement concentration and instrument status  
     When connecting via RS-232C interface, an RS-232C ↔ RS-485 converter should be used.

### Atmospheric pressure correction:

Measure atmospheric pressure and calculate compensation (for use, be sure to relieve the exhaust gas from analyzer to the atmosphere)

After atmospheric pressure correction;

Zero point: No influenced

Span point: The change is 0.5% measured value or less relating to the change of the atmospheric pressure 1%.

Correction range: 700hPa-1050hPa

## Performance

### Repeatability:

±0.5% of full scale

### Linearity:

1% of full scale

prior to atmospheric pressure correction (option)

### Zero drift:

±2% of full scale/week

In the case of NO and/or SO<sub>2</sub> measurement below 500 ppm range, with the auto zero calibration used.

### Span drift:

±2% of full scale/week

### Response time (for 90% FS response) :

1 to 15 sec electrical response. Within 10-30 seconds including replacement time of sampling gas.

Gas replacement time depends on the number of measuring components, and measuring range.

Interference from other gases:

Interference component	CO <sub>2</sub> analyzer	CO analyzer	CH <sub>4</sub> analyzer	SO <sub>2</sub> analyzer	NO analyzer
CO 1000 ppm	≤ 1.0 % FS	–	≤ 1.0 % FS	≤ 1.0 % FS	≤ 1.0 % FS
CO <sub>2</sub> 15%	–	≤ 1.0 % FS <sup>*1</sup>	≤ 1.0 % FS	≤ 1.0 % FS	≤ 1.0 % FS <sup>*2</sup>
H <sub>2</sub> O saturation at 20°C	≤ 1.0 % FS	≤ 1.0 % FS <sup>*3</sup>	≤ 1.0 % FS	–	–
H <sub>2</sub> O saturation at 2°C	–	≤ 2.0 % FS	–	≤ 2.0 % FS	≤ 2.0 % FS
CH <sub>4</sub> 1000 ppm	≤ 1.0 % FS	≤ 1.0 % FS	–	≤ 20 ppm	–

\*1: 0–200 ppm range: ≤ 2.0% FS

\*2: 0–500 ppm range: ≤ 2.0% FS

\*3: 0–500 ppm range: ≤ 2.0% FS

Interference for 0–200 ppm range may be greater than 2.0% FS depending on conditions.

## Requirements for Sample Gas

### Flow rate:

0.5 ±0.2L / min

### Temperature:

0 to 50°C

### Pressure:

10 kPa or less (Gas outlet side should be open to the atmospheric air.)

### Dust:

100 µg/Nm<sup>3</sup> or less in particle size of 0.3 µm or smaller

### Mist:

Unallowable

### Moisture:

For sample gases NO, SO<sub>2</sub>, CO (0-200 ppm range): less than 2°C saturation point.

For most other sample gases: less than standard room temperature saturation point.

### Corrosive component:

1 ppm or less

### Standard gas for calibration:

1) For measurement with IR and/or built-in O<sub>2</sub> sensor  
 Zero gas; Dry N<sub>2</sub>  
 Span gas; Each sample gas having concentration of 90 to 100% of its measuring range (recommended).

2) For measurement with external zirconia O<sub>2</sub> sensor and when calibration is carried out on the same calibration gas line:

Zero gas; Dry air or atmospheric air (This is not allowed for CO<sub>2</sub> measurement.)

Span gas; For other than O<sub>2</sub> measurement, each sample gas having concentration of 90 to 100% of its measuring range

\* For O<sub>2</sub> measurement, 1–2 vol% O<sub>2</sub>, balance N<sub>2</sub>

3) Reverse range O<sub>2</sub> measurement

Zero gas; 100vol% O<sub>2</sub>

Span gas; 95.0 to 95.5 vol% O<sub>2</sub>, balance N<sub>2</sub>

\* If you use the reverse range O<sub>2</sub> measurement, IR measurement is not available.

## Installation Requirements

- Indoor use. Select a place where the equipment does not receive direct sunlight, wind and rain, or radiation from hot substances. If such a place cannot be found, a roof or cover should be prepared for protection.
- Avoid a place where unit receives heavy vibration
- Select a place where atmospheric air is clean

## EU Directive Compliance

LVD (2014/35/EU)

EN 61010-1

EN 62311

EMC (2014/30/EU)

EN 61326-1(Table 2)

EN 61000-3-2(Class A)

EN 61000-3-3


EN61326-2-3

RoHS (2011/65/EU+(EU)2015/863)


EN IEC63000

**PRINCIPLE**


**NDIR sensor (for CO<sub>2</sub>, CO, CH<sub>4</sub>, SO<sub>2</sub>, NO)**



**Paramagnetic sensor (for O<sub>2</sub>)**




**Galvanic fule cell (for O<sub>2</sub>)**




# SYSTEM CONFIGURATION EXAMPLES

## CO, CO<sub>2</sub>, or CH<sub>4</sub> with low-level moisture



## NO, SO<sub>2</sub>, 0-200 ppm range CO, or other gases with high-level moisture



\*1) For NO, SO<sub>2</sub>, or 0–200 ppm range CO measurement, be sure to use an electronic cooler to keep the moisture content below the level saturated at 2°C.

## List of sampling devices (example)

No.	Device name	Fuji's type
①	Mist filter	ZBBK1V03-0
②	Safety drain trap	ZBH51603
③	Pump	ZBG80
④	Electronic cooler	ZBC91004
⑤	Drain pot	ZBH13003 (Length 255mm)
⑥	Ball valve	ZBFB1
⑦	Two-way solenoid valve	
⑧	Standard gas for calibration	ZBM□Y04-0□ (Codes in □ to be selected depending on application)
⑨	Flow meter	ZBD42203
⑩	Membrane filter	ZBBM2V03-0
⑪	Demister	ZBH35003
⑫	NO <sub>2</sub> /NO converter	ZDL02001
⑬	Three-way solenoid valve	

Note) The above is a typical configuration example. As configuration may differ depending on measuring objects, please consult us.



Digit	Description	note	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	← Digit
23	<Pressure compensation> None Pressure compensation		Z	P	A	B				2																			
24	<Unit> ppm, vol% mg/m <sup>3</sup> , g/m <sup>3</sup>	note8																											
25	<Adjustment> For standard For heat treatment furnace For converter Others	note9  note10																											
26	<Others> Non-standard																												Z

## NDIR range codes

Range	Code	Range	Code
None	Y	0 to 1vol%	J
0 to 100ppm	B	0 to 2vol%	K
0 to 200ppm	C	0 to 3vol%	Q
0 to 250ppm	D	0 to 5vol%	L
0 to 300ppm	S	0 to 10vol%	M
0 to 500ppm	E	0 to 20vol%	N
0 to 1000ppm	F	0 to 25vol%	V
0 to 2000ppm	G	0 to 40vol%	W
0 to 2500ppm	U	0 to 50vol%	P
0 to 3000ppm	T	0 to 70vol%	X
0 to 5000ppm	H	0 to 100vol%	R
		Others	Z

## O<sub>2</sub> range codes

Measurement range	Range code	Galvanic fuel cell (built - in)	Paramagnetic cell (built - in)	Zirconia cell (external)
0 to 5/10 vol%	A		○ note11	○
0 to 5/25 vol%	B		○ note11	○
0 to 10/25 vol%	C	○	○	○
0 to 5 vol%	L		○ note11	○
0 to 10 vol%	M	○	○	○
0 to 25 vol%	V	○	○	○
0 to 50 vol%	P		○	
0 to 100 vol%	R		○	
100 to 95 vol%	S		○	

- note1) If you select "D" in the 4th code, the analyzer comes with the power cable. Specify the rating of the power cable in the 20th code.
- note2) If you use this analyzer only for oxygen measurement, select "Y" in the 6th code.
- note3) If you use an external O<sub>2</sub> sensor (7th code "1"), set 0–1 V DC linear signals from the external O<sub>2</sub> analyzer so that they corresponds to the full scale setting of the analyzer.  
Note that the external O<sub>2</sub> analyzer (7th code "1") and the external zirconia O<sub>2</sub> sensor ZFK7 (7th code "2") need to be ordered separately.
- note4) Check the possible combination of measuring components and ranges in Table 1. Specify the range with the range codes shown in the above tables.
- note5) If you select "D" in the 4th code, select the appropriate cable specification for end user in the 20th code.  
If you select "A" in the 4th code, the power cable is not supplied. If you need no power cable and want to order the manual written in English, select "E" in the 20th code.
- note6) O<sub>2</sub> correction is provided only for NO, SO<sub>2</sub>, and CO measurement.  
When "H" is specified for the 6th digit, "A, B, or C" cannot be selected for the 21st digit.
- note7) The 22nd code "H" is not available for five-component measurement.  
If you use four-component measurement and select "H" in the 22nd code, the maximum number of the H/L alarm outputs is three.
- note8) Even if you selected "B" in the 24th code, select the range in ppm that is shown in the above "NDIR range codes" table. We will set the analyzer after converting the ppm ranges into mg/m<sup>3</sup> ranges. For the converted ranges, see the "ppm-mg/m<sup>3</sup> conversion table" shown below.
- note9) When A, C, or D is specified at 25th digit, the analyzer will be adjusted and delivered with the following balance gasses.  
Standard "A": balance gas N<sub>2</sub>  
For heat treatment furnace "C": CO<sub>2</sub> analyzer: 25% CO + 30% H<sub>2</sub> + residue N<sub>2</sub>  
CO analyzer: 5% CO<sub>2</sub> + 30% H<sub>2</sub> + residue N<sub>2</sub>  
CH<sub>4</sub> analyzer: 25% CO + 30% H<sub>2</sub> + residue N<sub>2</sub>  
For converter "D": balance gas CO, CO  
When other adjustment is required, please select "Z". In that case, please provide a list of gas composition of the process gas.
- note10) When the 25th code is "C", the range codes "X" and "R" are not available.
- note11) NDIR range codes "V", "W", "P", "X", and "R" are not available.

ppm–mg/m<sup>3</sup> conversion table

Range code	ppm	Range in mg/m <sup>3</sup>		
		NO	SO <sub>2</sub>	CO
C	0–200	—	—	0–250 mg/m <sup>3</sup>
D	0–250	—	—	0–300 mg/m <sup>3</sup>
S	0–300	—	—	0–375 mg/m <sup>3</sup>
E	0–500	0–650 mg/m <sup>3</sup>	0–1400 mg/m <sup>3</sup>	0–600 mg/m <sup>3</sup>
F	0–1000	0–1300 mg/m <sup>3</sup>	0–2800 mg/m <sup>3</sup>	0–1250 mg/m <sup>3</sup>
G	0–2000	0–2600 mg/m <sup>3</sup>	0–5600 mg/m <sup>3</sup>	0–2500 mg/m <sup>3</sup>
U	0–2500	0–3300 mg/m <sup>3</sup>	0–7100 mg/m <sup>3</sup>	0–3000 mg/m <sup>3</sup>
T	0–3000	0–4000 mg/m <sup>3</sup>	0–8500 mg/m <sup>3</sup>	0–3750 mg/m <sup>3</sup>
H	0–5000	0–6600 mg/m <sup>3</sup>	0–14.00 g/m <sup>3</sup>	0–6250 mg/m <sup>3</sup>

Conversion formula  
 NO (mg/m<sup>3</sup>) = 1.34 × NO (ppm)  
 SO<sub>2</sub> (mg/m<sup>3</sup>) = 2.86 × SO<sub>2</sub> (ppm)  
 CO (mg/m<sup>3</sup>) = 1.25 × CO (ppm)

**Table 1 Measurable component and range - availability check table -**

Range selection procedure

One component analyzer:

First determine 1st range, then select 2nd range from the corresponding column.

Two or more component analyzer:

1. Check the available options referring to the below tables, and determine the 1st range. Select the corresponding range code from "NDIR range codes" on Page 8.
2. The 2nd range must be greater than the 1st range but no greater than the value in the column "2nd range (max.)". Check the available options referring to the below tables, and determine the 2nd range. Select the corresponding range code from "NDIR range codes" on Page 8.

## 1-component analyzer : CO

1st range	2nd range
0 - 200ppm	None, 0 - 250ppm,300ppm,500ppm,1000ppm,2000ppm
0 - 250ppm	None, 0 - 300ppm,500ppm,1000ppm,2000ppm,2500ppm
0 - 300ppm	None, 0 - 500ppm,1000ppm,2000ppm,2500ppm
0 - 500ppm	None, 0 - 1000ppm,2000ppm,2500ppm,3000ppm,5000ppm
0 - 1000ppm	None, 0 - 2000ppm,2500ppm,3000ppm,5000ppm,1%
0 - 2000ppm	None, 0 - 2500ppm,3000ppm,5000ppm,1%,2%
0 - 2500ppm	None, 0 - 3000ppm,5000ppm,1%,2%
0 - 3000ppm	None, 0 - 5000ppm,1%,2%
0 - 5000ppm	None, 0 - 1%,2%,3%,5%
0 - 1%	None, 0 - 2%,3%,5%,10%
0 - 2%	None, 0 - 3%,5%,10%,20%
0 - 3%	None, 0 - 5%,10%,20%,25%
0 - 5%	None, 0 - 10%,20%,25%,40%,50%
0 - 10%	None, 0 - 20%,25%,40%,50%,70%,100%
0 - 20%	None, 0 - 25%,40%,50%,70%,100%
0 - 25%	None, 0 - 40%,50%,70%,100%
0 - 40%	None, 0 - 50%,70%,100%
0 - 50%	None, 0 - 70%,100%
0 - 70%	None, 0 - 100%
0 - 100%	None

1-component analyzer : CO<sub>2</sub>

1st range	2nd range
0 - 100ppm	None, 0 - 200ppm,250ppm,300ppm,500ppm,1000ppm
0 - 200ppm	None, 0 - 250ppm,300ppm,500ppm,1000ppm,2000ppm
0 - 250ppm	None, 0 - 300ppm,500ppm,1000ppm,2000ppm,2500ppm
0 - 300ppm	None, 0 - 500ppm,1000ppm,2000ppm,2500ppm
0 - 500ppm	None, 0 - 1000ppm,2000ppm,2500ppm,3000ppm,5000ppm
0 - 1000ppm	None, 0 - 2000ppm,2500ppm,3000ppm,5000ppm,1%
0 - 2000ppm	None, 0 - 2500ppm,3000ppm,5000ppm,1%,2%
0 - 2500ppm	None, 0 - 3000ppm,5000ppm,1%,2%
0 - 3000ppm	None, 0 - 5000ppm,1%,2%
0 - 5000ppm	None, 0 - 1%,2%,3%,5%
0 - 1%	None, 0 - 2%,3%,5%,10%
0 - 2%	None, 0 - 3%,5%,10%,20%
0 - 3%	None, 0 - 5%,10%,20%,25%
0 - 5%	None, 0 - 10%,20%,25%,40%,50%
0 - 10%	None, 0 - 20%,25%,40%,50%,70%,100%
0 - 20%	None, 0 - 25%,40%,50%,70%,100%
0 - 25%	None, 0 - 40%,50%,70%,100%
0 - 40%	None, 0 - 50%,70%,100%
0 - 50%	None, 0 - 70%,100%
0 - 70%	None, 0 - 100%
0 - 100%	None

## 1-component analyzer : NO

1st range	2nd range
0 - 200ppm	None, 0 - 250ppm,300ppm,500ppm,1000ppm,2000ppm
0 - 250ppm	None, 0 - 300ppm,500ppm,1000ppm,2000ppm,2500ppm
0 - 300ppm	None, 0 - 500ppm,1000ppm,2000ppm,2500ppm
0 - 500ppm	None, 0 - 1000ppm,2000ppm,2500ppm,3000ppm,5000ppm
0 - 1000ppm	None, 0 - 2000ppm,2500ppm,3000ppm,5000ppm
0 - 2000ppm	None, 0 - 2500ppm,3000ppm,5000ppm
0 - 2500ppm	None, 0 - 3000ppm,5000ppm
0 - 3000ppm	None, 0 - 5000ppm
0 - 5000ppm	None

1-component analyzer : SO<sub>2</sub>

1st range	2nd range
0 - 200ppm	None, 0 - 250ppm,300ppm,500ppm,1000ppm,2000ppm
0 - 250ppm	None, 0 - 300ppm,500ppm,1000ppm,2000ppm,2500ppm
0 - 300ppm	None, 0 - 500ppm,1000ppm,2000ppm,2500ppm
0 - 500ppm	None, 0 - 1000ppm,2000ppm,2500ppm,3000ppm,5000ppm
0 - 1000ppm	None, 0 - 2000ppm,2500ppm,3000ppm,5000ppm,1%
0 - 2000ppm	None, 0 - 2500ppm,3000ppm,5000ppm,1%,2%
0 - 2500ppm	None, 0 - 3000ppm,5000ppm,1%,2%
0 - 3000ppm	None, 0 - 5000ppm,1%,2%
0 - 5000ppm	None, 0 - 1%,2%,3%,5%
0 - 1%	None, 0 - 2%,3%,5%,10%
0 - 2%	None, 0 - 3%,5%,10%
0 - 3%	None, 0 - 10%
0 - 5%	None, 0 - 10%
0 - 10%	None

1-component analyzer : CH<sub>4</sub>

1st range	2nd range
0 - 500ppm	None, 0 - 1000ppm,2000ppm,2500ppm,3000ppm,5000ppm
0 - 1000ppm	None, 0 - 2000ppm,2500ppm,3000ppm,5000ppm,1%
0 - 2000ppm	None, 0 - 2500ppm,3000ppm,5000ppm,1%,2%
0 - 2500ppm	None, 0 - 3000ppm,5000ppm,1%,2%
0 - 3000ppm	None, 0 - 5000ppm,1%,2%
0 - 5000ppm	None, 0 - 1%,2%,3%,5%
0 - 1%	None, 0 - 2%,3%,5%,10%
0 - 2%	None, 0 - 3%,5%,10%,20%
0 - 3%	None, 0 - 5%,10%,20%,25%
0 - 5%	None, 0 - 10%,20%,25%,40%,50%
0 - 10%	None, 0 - 20%,25%,40%,50%,70%,100%
0 - 20%	None, 0 - 25%,40%,50%,70%,100%
0 - 25%	None, 0 - 40%,50%,70%,100%
0 - 40%	None, 0 - 50%,70%,100%
0 - 50%	None, 0 - 70%,100%
0 - 70%	None, 0 - 100%
0 - 100%	None





3-component analyzer: NO/SO<sub>2</sub>/CO >>> Combination of 1st component NO and 2nd component SO<sub>2</sub>/ 3rd component CO

1-component: NO		+	2-component: SO <sub>2</sub>		3-component: CO	
1st range	2nd range (max.)		1st range	2nd range (max.)	1st range/2nd range (max.)	
0-200ppm	0-2000ppm		0-200ppm	0-2000ppm	0-200/2000ppm, 0-250/2500ppm, 0-300/2500ppm, 0-500/2500ppm, 0-1000/2500ppm, 0-2000/2500ppm, 0-2500ppm	
0-250ppm	0-2500ppm		0-250ppm	0-2500ppm		
0-300ppm	0-2500ppm		0-300ppm	0-2500ppm		
0-500ppm	0-5000ppm		0-500ppm	0-2500ppm	0-200/2000ppm, 0-250/2500ppm, 0-300/2500ppm, 0-500/5000ppm, 0-1000/5000ppm, 0-2000/5000ppm, 0-5000ppm	
0-1000ppm	0-5000ppm		0-1000ppm	0-2500ppm		
0-2000ppm	0-5000ppm		0-2500ppm	None		
0-2500ppm	0-5000ppm		0-1000ppm	0-5000ppm	0-500/5000ppm, 0-1000/5000ppm, 0-2000/5000ppm, 0-2500/5000ppm, 0-3000/5000ppm, 0-5000ppm	
0-3000ppm	0-5000ppm		0-2000ppm			
0-5000ppm	None		0-2500ppm			
			0-3000ppm			
		0-5000ppm	None			

3-component analyzer: CO<sub>2</sub>/CO/CH<sub>4</sub> >>> Combination of 1st component CO<sub>2</sub> / 2nd component CO and 3rd component CH<sub>4</sub>

1-component: CO <sub>2</sub>		2-component: CO		+	3-component: CH <sub>4</sub>		Availability of product
1st range	2nd range (max.)	1st range/2nd range (max.)			1st range	2nd range (max.)	
0-5000ppm	0-3%	0-500/5000ppm, 0-1000ppm/1%, 0-2000ppm/2%, 0-2500ppm/2%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-5000ppm	0-5%	Product available only when CO analyzer max. measuring range is 50% or less
0-1%	0-5%				0-1%	0-10%	Product available
0-2%	0-5%				0-2%	0-20%	
0-5000ppm	0-5%	0-500/5000ppm, 0-1000ppm/1%, 0-2000ppm/2%, 0-2500ppm/2%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/50%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-3%	0-25%	Product available only when CO analyzer measuring range is 0 to 1000ppm or more.
0-1%	0-10%	0-500/5000ppm, 0-1000ppm/1%, 0-2000ppm/1%, 0-2500ppm/1%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/25%, 0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-5%	0-10%	Product available only when CO analyzer measuring range is 0 to 5000ppm or more.
0-2%	0-20%	0-500/5000ppm, 0-1000ppm/1%, 0-2000ppm/1%, 0-2500ppm/1%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/50%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-10%	0-20%	Available only when the CO analyzer range is 0-5000 ppm or more, and the CO <sub>2</sub> analyzer range is 0-2% or more
0-2%	0-10%	0-500/5000ppm, 0-1000ppm/1%, 0-2000ppm/1%, 0-2500ppm/1%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-20%	0-25%	
0-3%	0-25%	0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-25%	0-40%	
0-5%	0-50%	0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-40%	0-50%	
0-10%	0-100%	0-1000ppm/1%, 0-2000ppm/1%, 0-2500ppm/1%, 0-3000ppm/2%, 0-5000ppm/5%, 0-1/10%, 0-2/20%, 0-3/25%, 0-5/50%, 0-10/100%, 0-20/100%, 0-25/100%, 0-40/100%, 0-50/100%, 0-70/100%, 0-100%			0-50%	0-70%	
0-20%				0-70%	0-100%		
0-25%				0-100%	None		
0-40%							
0-50%							
0-70%							
0-100%	None						

4-component analyzer: NO/SO<sub>2</sub>/CO<sub>2</sub>/CO >>> Combination of 1st component NO /4th component CO and component 2nd component SO<sub>2</sub>/3rd component CO<sub>2</sub>

1-component: NO		4-component: CO	
1st range	2nd range (max.)	1st range/2nd range (max.)	
0-200ppm	0-2000ppm	0-200/2000ppm, 0-250/2500ppm, 0-300/2500ppm, 0-500/2500ppm, 0-1000/2500ppm, 0-2000/2500ppm, 0-2500ppm, None	
0-250ppm	0-2500ppm		
0-300ppm	0-2500ppm		
0-500ppm	0-2000ppm		
0-1000ppm	0-2000ppm		
0-2000ppm	None		
0-500ppm	0-5000ppm		
0-1000ppm	0-5000ppm		
0-2000ppm	0-5000ppm		
0-2500ppm	0-5000ppm		
0-3000ppm	0-5000ppm	0-500/2500ppm, 0-1000/2500ppm, 0-2000/2500ppm, 0-2500ppm, None	
0-5000ppm	None		




2-component analyzer: SO <sub>2</sub>		3-component analyzer: CO <sub>2</sub>	
1st range	2nd range (max.)	1st range/2nd range (max.)	
0-200ppm	0-2000ppm	0-1/10%, 0-2/20%, 0-3/20%, 0-5/50%, 0-10/50%, 0-20/50%, 0-25/50%, 0-40/50%, 0-50%/None	
0-250ppm	0-2500ppm		
0-300ppm	0-2500ppm		
0-500ppm	0-5000ppm		
0-1000ppm	0-5000ppm		
0-2000ppm	0-5000ppm		
0-2500ppm	0-5000ppm		
0-3000ppm	0-5000ppm		
0-5000ppm	None		

Table 2 Channel (Ch) No. and display/output contents comparison table

Code symbol			Display/output contents
6th digit	7th digit	21st digit	
Y	1 to 4	Y	Ch1:O <sub>2</sub>
P	Y	Y	Ch1:NO
A	Y	Y	Ch1:SO <sub>2</sub>
D	Y	Y	Ch1:CO <sub>2</sub>
B	Y	Y	Ch1:CO
E	Y	Y	Ch1:CH <sub>4</sub>
F	Y	Y	Ch1:NO, Ch2:SO <sub>2</sub>
G	Y	Y	Ch1:NO, Ch2:CO
H	Y	Y	Ch1:SO <sub>2</sub> , Ch2:CO <sub>2</sub>
J	Y	Y	Ch1:CO <sub>2</sub> , Ch2:CO
K	Y	Y	Ch1:CH <sub>4</sub> , Ch2:CO
L	Y	Y	Ch1:CO <sub>2</sub> , Ch2:CH <sub>4</sub>
N	Y	Y	Ch1:NO, Ch2:SO <sub>2</sub> , Ch3:CO
T	Y	Y	Ch1:CO <sub>2</sub> , Ch2:CO, Ch3:CH <sub>4</sub>
V	Y	Y	Ch1:NO, Ch2:SO <sub>2</sub> , Ch3:CO <sub>2</sub> , Ch4:CO
P	1 to 4	Y	Ch1:NO, Ch2:O <sub>2</sub>
A	1 to 4	Y	Ch1:SO <sub>2</sub> , Ch2:O <sub>2</sub>
D	1 to 4	Y	Ch1:CO <sub>2</sub> , Ch2:O <sub>2</sub>
B	1 to 4	Y	Ch1:CO, Ch2:O <sub>2</sub>
E	1 to 4	Y	Ch1:CH <sub>4</sub> , Ch2:O <sub>2</sub>
F	1 to 4	Y	Ch1:NO, Ch2:SO <sub>2</sub> , Ch3:O <sub>2</sub>
G	1 to 4	Y	Ch1:NO, Ch2:CO, Ch3:O <sub>2</sub>
H	1 to 4	Y	Ch1:SO <sub>2</sub> , Ch2:CO <sub>2</sub> , Ch3:O <sub>2</sub>
J	1 to 4	Y	Ch1:CO <sub>2</sub> , Ch2:CO, Ch3:O <sub>2</sub>
K	1 to 4	Y	Ch1:CH <sub>4</sub> , Ch2:CO, Ch3:O <sub>2</sub>
L	1 to 4	Y	Ch1:CO <sub>2</sub> , Ch2:CH <sub>4</sub> , Ch3:O <sub>2</sub>
N	1 to 4	Y	Ch1:NO, Ch2:SO <sub>2</sub> , Ch3:CO, Ch4:O <sub>2</sub>
T	1 to 4	Y *	Ch1:CO <sub>2</sub> , Ch2:CO, Ch3:CH <sub>4</sub> , Ch4:O <sub>2</sub>
V	1 to 4	Y *	Ch1:NO, Ch2:SO <sub>2</sub> , Ch3:CO <sub>2</sub> , Ch4:CO, Ch5:O <sub>2</sub>
P	1 to 4	A *	Ch1:NO <sub>x</sub> , Ch2:O <sub>2</sub> , Ch3:corrected NO <sub>x</sub>
A	1 to 4	A *	Ch1:SO <sub>2</sub> , Ch2:O <sub>2</sub> , Ch3:corrected SO <sub>2</sub>
B	1 to 4	A *	Ch1:CO, Ch2:O <sub>2</sub> , Ch3:corrected CO
F	1 to 4	A *	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:O <sub>2</sub> , Ch4:corrected NO <sub>x</sub> , Ch5:corrected SO <sub>2</sub>
G	1 to 4	A *	Ch1:NO <sub>x</sub> , Ch2:CO, Ch3:O <sub>2</sub> , Ch4:corrected NO <sub>x</sub> , Ch5:corrected CO
J	1 to 4	A *	Ch1:CO <sub>2</sub> , Ch2:CO, Ch3:O <sub>2</sub> , Ch4:corrected CO
N	1 to 4	A *	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:CO, Ch4:O <sub>2</sub> , Ch5:corrected NO <sub>x</sub> , Ch6:corrected SO <sub>2</sub> , Ch7:corrected CO
V	1 to 4	A *	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:CO <sub>2</sub> , Ch4:CO, Ch5:O <sub>2</sub> , Ch6:corrected NO <sub>x</sub> , Ch7:corrected SO <sub>2</sub> , Ch8:corrected CO
P	1 to 4	C *	Ch1:NO <sub>x</sub> , Ch2:O <sub>2</sub> , Ch3:corrected NO <sub>x</sub> , Ch4:corrected NO <sub>x</sub> average
A	1 to 4	C *	Ch1:SO <sub>2</sub> , Ch2:O <sub>2</sub> , Ch3:corrected SO <sub>2</sub> , Ch4:corrected SO <sub>2</sub> average
B	1 to 4	C	Ch1:CO, Ch2:O <sub>2</sub> , Ch3:corrected CO, Ch4:corrected CO average
F	1 to 4	C *	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:O <sub>2</sub> , Ch4:corrected NO <sub>x</sub> , Ch5:corrected SO <sub>2</sub> , Ch6:corrected NO <sub>x</sub> average, Ch7:corrected SO <sub>2</sub> average
G	1 to 4	C *	Ch1:NO <sub>x</sub> , Ch2:CO, Ch3:O <sub>2</sub> , Ch4:corrected NO <sub>x</sub> , Ch5:corrected CO, Ch6:corrected NO <sub>x</sub> average, Ch7:corrected CO average
J	1 to 4	C	Ch1:CO <sub>2</sub> , Ch2:CO, Ch3:O <sub>2</sub> , Ch4:corrected CO, Ch5:corrected CO average
N	1 to 4	C *	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:CO, Ch4:O <sub>2</sub> , Ch5:corrected NO <sub>x</sub> , Ch6:corrected SO <sub>2</sub> , Ch7:corrected CO, Ch8:corrected NO <sub>x</sub> average, Ch9:corrected SO <sub>2</sub> average, Ch10:corrected CO average
V	1 to 4	C	Ch1:NO <sub>x</sub> , Ch2:SO <sub>2</sub> , Ch3:CO <sub>2</sub> , Ch4:CO, Ch5:O <sub>2</sub> , Ch6:corrected NO <sub>x</sub> , Ch7:corrected SO <sub>2</sub> , Ch8:corrected CO, Ch9:corrected NO <sub>x</sub> average, Ch10:corrected SO <sub>2</sub> average, Ch11:corrected CO average

\* When the 21st digit code is A or C, the component of the NO analyzer is displayed as NO<sub>x</sub>.


# DIMENSIONS (Unit : mm)





## Mounting method

The analyzer weight should be supported at the bottom of the case.

### 19-inch rack mounting type



### Mounting diagram



## SCOPE OF DELIVERY


- Gas analyzer ... 1 unit
- Replacement fuse (250V AC, 2A, delay type) ... 2 pcs
- Instruction manual ... 1 copy
- Connector for I/O connection ... 1 set
- When the 4th code is "D": power supply cord (standard inlet type 2m) ... 1 pc

## ORDERING INFORMATION

1. Code symbols
2. Application and composition of sample gas

## EXTERNAL CONNECTION


### <Analog output> A/O connector





D-sub 25pins female

\* In standard, displayed Channel No. and Analog Output No. are same.


### <Screw terminal (M4)>




### <Power inlet>




### <RS485 communication signal>




D-sub 9pins female



### <Analog input> A/I connector (O<sub>2</sub> signal input)



<Digital I/O> DIO 1 to 3 connector (option)




D-sub 25pins female

\* DIO 1 to 3 are all same type of connector.

Contents of digital input signal

DI1	Remote hold
DI2	Average value reset
DI3	A. cal. start
DI4	A. zero. cal. start
DI5	Remote range Ch1
DI6	Remote range Ch2
DI7	Remote range Ch3
DI8	Remote range Ch4
DI9	Remote range Ch5



Digital input  
OFF : 0V  
ON : 12 to 24V DC

Contact capacity  
24V DC/1A

Allocation table of digital input signal

22th digit →	A	B	C	D	E	F	G	H	Y
DI1	○	○	○	○	○	○	○	○	
DI2	○	○	○	○	○	○	○	○	
DI3		○			○		○	○	
DI4		○			○		○	○	
DI5				○		○	○	○	
DI6				○*		○*	○*	○*	
DI7				○*		○*	○*	○*	
DI8				○*		○*	○*	○*	
DI9				○*		○*	○*	○*	

○ sign shows the function is valid.

\* : The function might be invalid depending on the number of measurable components.

For example: DI5 corresponds to 1st component, DI6 corresponds to 2nd components.

Contents of digital output signal

22th digit →	Independent on the number of component	1-component analyzer		2-component analyzer	3-component analyzer
		B, E	D, F, G, H	B, D, E, F, G, H	B, D, E, F, G, H
DO1	Instrument error	Instrument error	Instrument error	Instrument error	Instrument error
DO2	Calibration error	Calibration error	Calibration error	Calibration error	Calibration error
DO3		A.cal.status	(A.cal.status)	(A.cal.status)	(A.cal.status)
DO4		For zero gas	(For zero gas)	(For zero gas)	(For zero gas)
DO5		For span gas Ch1	(For span gas Ch1)	(For span gas Ch1)	(For span gas Ch1)
DO6	(Alarm1)	(Alarm1)		(For span gas Ch2)	(For span gas Ch2)
DO7	(Alarm2)	(Alarm2)			(For span gas Ch3)
DO8	(Alarm3)	(Alarm3)			(Range identification Ch1)
DO9	(Alarm4)	(Alarm4)		(Range identification Ch1)	(Range identification Ch2)
DO10	(Alarm5)	(Alarm5)	Range identification Ch1	(Range identification Ch2)	(Range identification Ch3)
DO11			(Alarm1)	(Alarm1)	(Alarm1)
DO12			(Alarm2)	(Alarm2)	(Alarm2)
DO13			(Alarm3)	(Alarm3)	(Alarm3)
DO14			(Alarm4)	(Alarm4)	(Alarm4)
DO15			(Alarm5)	(Alarm5)	(Alarm5)

The items in the parentheses may not be available depending on the selected type on 22th digit.

The normal open side (NO) of digital output is close when the function is active without range ID.

In case of range ID, normal open (NO) side is close with First range.

The normal close (NC) side is close with Second range.

22th digit →	4-component analyzer				5-component analyzer		
	B, E	D, F	G	H	B, E	D, F	G
DO1	Instrument error	Instrument error	Instrument error	Instrument error	Instrument error	Instrument error	Instrument error
DO2	Calibration error	Calibration error	Calibration error	Calibration error	Calibration error	Calibration error	Calibration error
DO3	A.cal.status		A.cal.status	A.cal.status	A.cal.status		A.cal.status
DO4	For zero gas		For zero gas	For zero gas	For zero gas		For zero gas
DO5	For span gas Ch1		For span gas Ch1	For span gas Ch1	For span gas Ch1		For span gas Ch1
DO6	For span gas Ch2		For span gas Ch2	For span gas Ch2	For span gas Ch2	Range identification Ch1	For span gas Ch2
DO7	For span gas Ch3	Range identification Ch1	For span gas Ch3	For span gas Ch3	For span gas Ch3	Range identification Ch2	For span gas Ch3
DO8	For span gas Ch4	Range identification Ch2	For span gas Ch4	For span gas Ch4	For span gas Ch4	Range identification Ch3	For span gas Ch4
DO9		Range identification Ch3		Range identification Ch1	For span gas Ch5	Range identification Ch4	For span gas Ch5
DO10		Range identification Ch4		Range identification Ch2		Range identification Ch5	
DO11	(Alarm1)	(Alarm1)		(Alarm1)	(Alarm1)	(Alarm1)	Range identification Ch1
DO12	(Alarm2)	(Alarm2)	Range identification Ch1	(Alarm2)	(Alarm2)	(Alarm2)	Range identification Ch2
DO13	(Alarm3)	(Alarm3)	Range identification Ch2	(Alarm3)	(Alarm3)	(Alarm3)	Range identification Ch3
DO14	(Alarm4)	(Alarm4)	Range identification Ch3	Range identification Ch3	(Alarm4)	(Alarm4)	Range identification Ch4
DO15	(Alarm5)	(Alarm5)	Range identification Ch4	Range identification Ch4	(Alarm5)	(Alarm5)	Range identification Ch5

# Zirconia O<sub>2</sub> analyzer (to be purchased separately)

**Measuring method:**

Zirconia system

**Measurable component and measuring range:**

Measurable component		Range
O <sub>2</sub>	Oxygen	0 to 25vol%

- Repeatability:** Within ± 0.5% of full scale
- Linearity:** Within ± 1% of full scale
- Zero drift:** Within ± 1% of full scale/week
- Span drift:** Within ± 2% of full scale/week
- Response time:** Approx. 20 seconds (for 90% response)
- Measured gas flow rate:** 0.5 ± 0.25L / min

**Notes:**

- If process gas is combustible, measurement error may occur due to oxygen contained in the process gas.
- If process gas is corrosive (for example, SO<sub>2</sub> beyond 250 ppm range), the service life of zirconia sensor may be shortened.

**Gas inlet/outlet size:**

Rc1/4 or NPT1/4

**Power supply:**

- Rated voltage; 100 to 115V AC or 200 to 240V AC
- Rated frequency; 50Hz/60Hz
- Max. rated power; 215VA (at start up)  
65VA (during normal operation)

**Enclosure:** Steel casing, for indoor application

**Indication:** Temperature indication (LED)

**Temperature alarm output:**

- Contact output 1 from A contact,
- Contact capacity 220V AC, 1A (resistive load)

**Outer dimensions (H x W x D):**

141 x 170 x 190mm


**Weight:** Approx. 3kg

**Finish color:** Munsell 5Y 7/1


## CODE SYMBOLS

Description														
1	2	3	4	5	6	7	8	9	10	11	12	13	Measuring method Zirconia method	
Z	F	K	7	Y	Y	4		Y	0	Y	Y			
7YY								Power supply 100 to 115V AC 50/60Hz(Standard) 200 to 240V AC 50/60Hz(Standard) 200 to 240V AC 50/60Hz(CE mark)	9					
									B					
									C					
								1	Gas inlet/outlet size					
									8	Rc 1/4 NPT 1/4				

## DIMENSIONS (Unit: mm)



## EXTERNAL CONNECTION DIAGRAM



Information in this catalog is subject to change without notice.  
Read the instruction manuals thoroughly before using the products.



Gate City Ohsaki, East Tower, 11-2, Osaki 1-chome, Shinagawa-ku, Tokyo 141-0032, Japan  
Phone: +81-3-5435-7111  
www.fujielectric.com  
www.fujielectric.com/products/instruments/